VERIFICATION OF CONFORMITY

S/N: 022870

No.: **ICR/VC/CUK260403**

Name and address of Applicant	PE «SheMax Company» 168, Myru str., Kontsovo, Uzhgorod district, Transcarpathian region, 89423, Ukraine
Name and address of manufacturer:	PE «SheMax Company» 168, Myru str., Kontsovo, Uzhgorod district, Transcarpathian region, 89423, Ukraine
Product name:	BIOSTOP PRO Professional Sterilization System
Product types:	B5, B10, B15, B20, B25, B30
Product trademark:	SheMax

Verification was carried within following scope:

Information on the Declaration of Conformity:

Result:	**Legislation:**	**Standard:**
✓	LVD [2014/35/EU]	EN 61010-1:2010+A1:2019 EN 61010-2-040:2015
✓	EMC [2014/30/EU]	EN IEC 61326-1:2021 EN IEC 61000-3-2:2019 EN 61000-3-3:2013
✓	RoHS [2011/65/EU]	EN IEC 63000:2018

The assessment process has been carried out in accordance with individual rules and conditions agreed with the applicant. Evaluation has been carried out in accordance with:

Test report:	№ 0403010/26	№ Tr2026 – 17/04/1
Tests conducted by:	LLC "TESTING CERTIFICATION CENTER "PIVDENTEST"	Limited Liability Company «Testing Center «Vyprobuvach»
Issue date:	17.04.2026	
Expiration date:	16.04.2031	

Remarks:

- VoC was issued on voluntary basis and does not imply meeting all essential requirements listed in Declaration of Conformity.
- For introducing this product on European market may be needed EC/EU-type examination conducted by appropriate Notified Body.

CEO, ICR Co., Ltd.

ICR Co. Ltd.
www.icrqa.com
www.icrpolska.com
cert@icrqa.com